

Product Technical Information

Effective: December 5, 2014

Re: Nickel Content of Agee Digit Widget® System, Item No DWD-232

Skeletal pins contained in the Digit Widget System are made of type 316, implant grade, Stainless steel. This alloy is composed of approximately 14% Nickel, among other metals.

For patients who are allergic to nickel, Digit Widget skeletal pins may not be suitable. We do not provide, or know of, an alternative pin for patients who are allergic to nickel.

Contact Hand Biomechanics Lab at 800-522-5778 if you need further technical information on the Digit Widget System.