

Product Technical Information

Effective: October 20, 2014

Re: Latex Content of Agee WristJack[®] System, Item No. CFD-147

The WristJack System device, instrumentation and packaging are ‘not made with natural rubber latex’.

Contact Hand Biomechanics Lab, Inc. at 800-522-5778 if you need further technical information on the WristJack System.