

Product Technical Information

Effective: March 1, 2019

Re: Nickel Content of PIP Fix™ System, Item No. PFX-426

Skeletal pins contained in the PIP Fix System are made of type 316, implant grade, stainless steel. This alloy is composed of approximately 14% nickel, among other metals.

For patients who are allergic to nickel, PIP Fix skeletal pins may not be suitable. We do not provide, or know of, an alternative pin for patients who are allergic to nickel.

Contact Hand Biomechanics Lab at 800-522-5778 if you need further technical information on the PIP Fix System.